

**HOSPITAL SUNGAI BULOH  
SELANGOR**

**Performance of AHP Rapid HBsAg Test**

**Intended use**

The AHP HBsAg Test cassette is a rapid lateral flow immunochromatographic assay designed for qualitative determination of hepatitis B surface antigen (HBsAg) in Whole Blood, Serum or Plasma. It is intended to be used by professional as a screening test and as an aid in the diagnosis of hepatitis B virus infection. Any reactive specimen with Rapid HBsAg Test must be confirmed with alternative testing method(s) and clinical findings.

**Manufacturer**

Allen Helathcare Products (M) Sdn Bhd. No. 17 Lrg Nagasari 2, Tmn NAgasari, 13600, Perai, Penang, Malaysia.

**Test Principle**

The AHP Rapid HBsAg Test Device (Whole Blood/Serum/Plasma) is a qualitative immunoassay for the detection of HBsAg in whole blood, serum or plasma. The membrane is pre-coated with anti-HBAG antibodies on the test line region of the Device. During testing, the whole blood, serum or plasma specimen reacts with the particle coated with anti-HBAG antibodies. The mixture migrates upward on the membrane chromatographically by capillary action to react with anti-HBAG antibodies on the membrane and generate a colored line. The presence of this colored line in the test region indicates a positive result, while its absence indicates a negative result. To serve as a procedural control, a colored line will always appear in the control line region indicating that proper volume of specimen has been added and membrane wicking has occurred.

**Test Kit**

Lot Number : HBA24057733WWB  
Expiry Date : 1/11/2025  
Date of Test : 23/8/2024

**Instrument Used**

Manually performed

**Reagent and Sample Preparation, Result Interpretation**

Refer to product package insert in the attachment

**Sample Used**

Sample Type = Serum  
Known Positive Sample = 10  
Known Negative Sample = 10  
Total samples used for analysis = 20



**Performance Analysis**

		Intended Results		
		Positive	Negative	Total
AHP RAPID HBSAG TEST	Positive	10	0	10
	Negative	0	10	10
	Total	10	10	20

Estimated sensitivity	100%
Estimated specificity	100%
Positive Predictive Value (PPV)	100%
Negative Predictive Value (NPV)	100%
Total Accuracy	100%
Imprecision, Coefficient of unalikeability (%CU)	0%

**Conclusion**

All samples used for this kit evaluation were previously tested using the chemiluminescence immunoassay (CLIA) method. The estimated sensitivity and specificity results as per manufacturer's claims. The imprecision, calculated as % CU, is less than 10%. Therefore, it is acceptable for use in point-of-care testing (POCT).

Prepared by	Reviewed by
	
<p>KHAIRUL NAZRIN BIN MOHD NASORI                      PEGAWAI SAINS (MIKROBIOLOGI)                      JABATAN PATOLOGI                      HOSPITAL SUNGAI BULOH</p>	<p>DR NUR HANANI BINTI AHMAD                      Pakar Patologi (Mikrobiologi Perubatan) UDG†                      MMC 42539 NSR 133854                      Jabatan Patologi                      Hospital Sg Buloh</p> <p>2/10/2024</p>

**Disclaimer:**

The study was conducted for the method verification of a POCT kit to be used exclusively at Hospital Sungai Buloh. The content of this report does not imply that this product is endorsed or recommended by the Hospital Sungai Buloh. This report may not be reviewed, abstracted, quoted, reproduced or translated, in part or in whole, without the prior written permission of Hospital Sungai Buloh. Hospital Sungai Buloh does not allow the use of this report or reference in any manner in the labeling or advertising of this product or kit package.

**HOSPITAL SUNGAI BULOH  
SELANGOR**

Performance of AHP Rapid HCV Test

**Intended use**

The AHP Rapid HCV Test cassette is a rapid lateral flow immunochromatographic assay designed for qualitative detection of antibodies to hepatitis C virus (HCV) in Whole Blood, Serum or Plasma. It is a qualitative immunoassay and considered as an initial screening test for detection of HCV antibodies. Any reactive specimen with the Rapid HCV Test must be confirmed with alternative testing method(s) and clinical findings. The test is for a professional use only.

**Manufacturer**

Allen Helathcare Products (M) Sdn Bhd. No. 17 Lrg Nagasari 2, Tmn NAgasari, 13600, Perai, Penang, Malaysia.

**Test Principle**

Rapid HCV Test is a chromatographic immunoassay (CIA) for the detection of antibodies to HCV in human serum/ plasma or whole blood. The Rapid HCV Test employs a unique recombinant binding protein which is conjugated to colloidal gold dye particles and antigen which are bound to the solid phase membrane. The venous or capillary (fingerstick) whole blood, serum or plasma is applied to the Sample window of test device. Buffer solution is provided in a buffer bottle. The Buffer facilitates the lateral flow of the specimen and test reagents and promotes the binding of the antibodies to the antigen. The specimen/buffer mixture migrates along the test strip by capillary action, reconstituting the conjugate. If present, the antibodies bind to the colloidal gold conjugated antibody binding protein. In a reactive sample, the dye conjugated-immune complex migrates on the nitrocellulose membrane and is captured by the antigens immobilized in the TEST (T) area producing a pink/purple line. In the absence of HCV antibodies, there is no pink/purple line in the TEST (T) area. The sample continues to migrate along the membrane and produces a pink/purple line in the CONTROL (C) area containing goat anti-mouse immunoglobulin G. This procedural control serves to demonstrate that specimen and reagents have been properly applied and have migrated through the device.

**Test Kit**

Lot Number	:	HCV24077927AWB
Expiry Date	:	01/01/2026
Date of Test	:	23/8/2024

**Instrument Used**

Manually performed

**Reagent and Sample Preparation, Result Interpretation**

Refer to product package insert in the attachment

**Sample Used**

Sample Type = Serum  
 Known Positive Sample = 10  
 Known Negative Sample = 10  
 Total samples used for analysis = 20

**Performance Analysis**

		Intended Result		
		Positive	Negative	Total
AHP RAPID HCV TEST	Positive	9	0	9
	Negative	1	10	11
	Total	10	10	20

Estimated sensitivity	90%
Estimated specificity	100%
Positive Predictive Value (PPV)	100%
Negative Predictive Value (NPV)	91%
Total Accuracy	95%
Imprecision, Coefficient of unalikeability (%CU)	8%

**Conclusion**

All samples used for this kit evaluation were previously tested using the chemiluminescence immunoassay (CLIA) method. The estimated sensitivity result is below the manufacturer's claims; however, it is still acceptable. The estimated specificity result aligns with the manufacturer's claim. The imprecision, calculated as % CU, is less than 10%. The kit is acceptable for use in point-of-care testing (POCT).

Prepared by	Reviewed by
	
KHAIRUL NAZRIN BIN MOHD NASORI PEGAWAI SAINS (MIKROBIOLOGI) JABATAN PATOLOGI HOSPITAL SUNGAI BULOH	DR NUR HANANI BINTI AHMAD Pakar Patologi (Mikrobiologi Perubatan) (M.D.) MMC 42539 NSR 133854 Jabatan Patologi Hospital Sg Buloh 2/10/2024

**Disclaimer:**

The study was conducted for the method verification of a POCT kit to be used exclusively at Hospital Sungai Buloh. The content of this report does not imply that this product is endorsed or recommended by the Hospital Sungai Buloh. This report may not be reviewed, abstracted, quoted, reproduced or translated, in part or in whole, without the prior written permission of Hospital Sungai Buloh. Hospital Sungai Buloh does not allow the use of this report or reference in any manner in the labeling or advertising of this product or kit package.



**NATIONAL PUBLIC HEALTH LABORATORY (NPHL)  
MINISTRY OF HEALTH MALAYSIA  
EVALUATION TEST REPORT OF SYPHILLIS RAPID TEST**

Product Name: AHP Rapid Syphilis (TP) Test

**Kit/ Assay Intended Used**

In-Vitro Diagnostic.

**Manufacturer**

Allen Healthcare Products (M) Sdn. Bhd.

**Test Principle**

The principle used in this kit The AHP Rapid Syphilis (TP) test is a qualitative test for the detection of antibodies to T. pallidum in human serum/plasma or whole blood. The Rapid Syphilis Test is a chromatographic immunoassay (CIA) for the detection of all antibodies, including IgM, IgG and IgA to T. pallidum in human serum, plasma or whole blood. T. pallidum specific antigens are precoated onto membrane as a capture reagent on the test region. During the test, specimen is allowed to react with the colloidal gold particles, which have been labeled with T. pallidum specific antigens, as well. Antibodies to T. pallidum, if present, a red colored band will develop on the membrane. Absence of this red colored band in the test region suggests a negative result. To serve as a procedural control, a red colored band in the control region will always appear regardless the presence of antibodies to T. pallidum. The results are observed in 10 to 20 minutes.

**Test Kit/ Assay Information**

Batch mp./Identifier:	TPV02CSWB	Lot Number:	TP24037533LWB
Manufacturing Date:	2024-02-05	Expiry Date:	2024-09-01

**Instrument Used**

Timer

**Reagent and sample Preparation**

Kindly refer to product package insert in the attachment

**Test/ Assay Used:**

Sample Type = Whole Blood

**Performance Analysis**

AHP Rapid Syphilis (TP) Test - Syphilis RTK	Known Sample (NPHL)		Total
	Positive	Negative	
Positive	50	0	50
Negative	0	50	50
<b>Total</b>	50	50	100

Positive Predictive Value: 100.00 %

Negative Predictive Value: 100.00 %

**Comments**

During the kit evaluation, we observed these findings: 1. The single-use buffer provided in this kit is placed in a plastic dropper which looks exactly like the plastic dropper used for collecting blood, serum or plasma samples. The buffer was not labelled to distinguish this difference. 2. The single-use buffer provided in this kit is placed in a plastic dropper with a sealed end and required the use of a scissors to cut of the end before dispensing the buffer solution. Possibility of reagent contamination may occur when users reuse the same scissors to cut off the ends of this buffer droppers without proper decontamination between each kit used.

<b>Prepared by</b>	<b>Reviewed/ Approved By</b>
Hannah Phoon	R.Pusparani

HANNAH PHOON YIK PHING	R.PUSPARANI A/P RAMASAMY
Evaluator/ Officer in Charge	Head of Bacteriology Section/ Unit, NPHL

Date: 2024-05-24

*Disclaimer: The institution declares no conflict of interest. The content of this report does not imply that this product is endorsed or recommended by the institution. This report may not be reviewed, abstracted, quoted, translated or reproduced without prior return consent of National Public Health Laboratory (NPHL), Ministry of Health Malaysia. NPHL does not allow the use of this report in any manner in the labelling of this product or kit package.*

**FOR REFERENCE ONLY  
NOT FOR OFFICIAL SUBMISSION**